



Setting Standards  
Across Seven Sectors

## E.E.C Declaration of Conformity

Steroplast Healthcare Ltd hereby declare that the devices described hereafter:

Product	Description	Steroplast Code
Sterochef Plasters	Assorted Box of 100	5629F

Have been classified as Class Is (Annex IX Rule 4) and are in conformity with the essential requirements and provisions of Council Directive 93/42/EEC as amended by Directive 2007/47/EC and are in conformity with the national standards transposing harmonised standards EN ISO 14971:2012, BS EN ISO 15223-1:2016, ISO 11135-1:2014 and EN ISO 13485:2016.

and are subject to the procedure set out in Annex V of the Directive 93/42/EEC as amended by Directive 2007/47/EC under the supervision of Notified Body Number 1639, SGS Belgium NV, SGS House Noorderlaan 87 2030 Antwerp Belgium. Certificate Number GB19/964596.

Steroplast Healthcare Ltd, Alpha Point, Bradnor Road, Manchester, M22 4TE  
T: 0044 (0)161 902 3030, F: 0044 (0)161 902 3040, e-mail: [sales@steroplast.co.uk](mailto:sales@steroplast.co.uk)

**EC Representative:** Advena Limited, Tower Business Centre, Swatar, BKR 4013 Malta

Head of Purchasing – Dawn McCarthy

Signed (on behalf of Operations Director Afzal Ramzan) –

Date – 14/07/2021

Issue 2

Steroplast Healthcare Ltd.  
Alpha Point  
Bradnor Road  
Manchester  
M22 4TE

Tel: 0161 902 3030  
Fax: 0161 902 3040

Steroplast Healthcare Limited, Alpha Point, Bradnor Road, Manchester M22 4TE England  
Tel: 0161 902 3030 Fax: 0161 902 3040 Email: [enquiries@steroplast.co.uk](mailto:enquiries@steroplast.co.uk) Web: [www.steroplast.co.uk](http://www.steroplast.co.uk)  
Directors: AC Brown (MD), A Ramzan, J Williams Registered No: 504340



Setting Standards Across Seven Sectors

Emergency

Hospital

Sport

Veterinary

Medical

First Aid

Retail